



Society for Maternal-Fetal Medicine Special Statement: Reducing the risk of transmitting infection by transvaginal ultrasound examination

SMFM Patient Safety and Quality Committee; Rebecca F. Hamm, MD; C. Andrew Combs, MD, PhD;
and Christina M. Davidson, MD

Recent surveys have shown widespread lapses in the procedures used to reduce the risk of transmitting infection via medical devices. Transvaginal ultrasound examination has the potential to transmit vaginal infections, including human papillomavirus. Areas of particular concern are the use of probe covers with high rates of leakage, disinfectants that are not effective against human papillomavirus, and coupling gel from multiple-use containers. We reviewed these issues, and we recommend 4 steps to reduce the risk of transmitting infection. First, during every transvaginal ultrasound exam, the probe should be covered with a sterile, single-use “viral barrier” cover or a condom. Second, sterile, single-use ultrasound gel packets should be used. Third, after every examination, the probe should be cleaned to remove any visible gel or debris. Finally, after cleaning, the probe should undergo high-level disinfection using an agent with proven efficacy against the human papillomavirus, including hydrogen peroxide, hypochlorite, or peracetic acid. Glutaraldehyde, orthophthalaldehyde, phenols, and isopropyl alcohol have virtually no efficacy against the human papillomavirus.

Key words: Glutaraldehyde, hydrogen peroxide, infection prevention, orthophthalaldehyde, ultrasound gel, ultrasound safety

Introduction

Transvaginal ultrasound examination has the potential to transmit vaginal infections between patients unless certain well-described procedures are followed to clean and disinfect the intravaginal ultrasound probe between examinations.¹ Two recent surveys have found widespread and extensive lapses in infection prevention practices related to ultrasound probes and other medical devices.^{2,3} The incidence of such lapses has been increasing in recent years.²

The purpose of this document is to review the recommended procedures to minimize the risk of transmitting infection via transvaginal ultrasound examination, with emphasis on recent data demonstrating that some disinfectants do not neutralize human papillomavirus (HPV).

Why is the disinfection of transvaginal probes an important issue?

Transvaginal ultrasound has become a routine part of obstetrical care. In the first trimester, transvaginal imaging is a key adjunct to transabdominal imaging for evaluating

threatened pregnancy loss, ectopic pregnancy, and fetal anomalies.^{4,5} In second-trimester screening for the risk of preterm birth, transvaginal measurement of cervical length is superior to transabdominal measurement.^{6,7} Some experts have advocated for universal transvaginal cervical length screening for all singleton pregnancies,^{8–11} although universal screening has not been mandated by the Society for Maternal-Fetal Medicine.¹²

Clinically significant pathogenic bacteria and viruses can survive on transvaginal probes. These include strains of *Staphylococcus*, *Bacillus*, *Pseudomonas*, *Enterococcus*, *Clostridium*, *Chlamydia*, and HPV.^{13–15} Of particular concern, HPV types 16 and 18 can withstand harsh conditions, including several widely used disinfectant methods.^{16,17} These high-risk HPV strains are the most common causes of cervical cancer.^{18,19}

Improper cleaning and disinfection of ultrasound probes are clear problems that require urgent attention in the United States. In a 2016 survey by The Joint Commission, 74% of all discovered immediate threats to life were related to improperly sterilized or disinfected equipment.² The rate of noncompliance with The Joint Commission infection prevention standards was 60% in hospitals, 57% in office-based settings, and 53% in ambulatory healthcare

Corresponding author: Patient Safety and Quality Committee, Society for Maternal-Fetal Medicine. smfm@smfm.org

BOX**Reasons for noncompliance with infection prevention standards in surveys by The Joint Commission**

Mistaken belief that the risk of transmitting pathogens to patients is low or nonexistent.
 Staff lacks the knowledge or training required to disinfect equipment properly.
 Staff does not have access to or lacks knowledge of evidence-based guidelines.
 Leadership oversight is lacking.
 Disinfection of equipment is a low priority within the organization.
 There is a lack of a culture of safety that supports the reporting of safety risks.
 Processes for disinfection are not followed (ie, staff take shortcuts).
 The time frames for proper disinfection of equipment are not followed.
 There is no dedicated staff member to oversee the proper disinfection of equipment.
 Facility design or space issues prevent proper disinfection of equipment (eg, processing takes place in a small room that is also used for storage).
 There is a lack of monitoring or documentation of disinfection of equipment, which makes it difficult to track the use of equipment on a specific patient, complicating the patient notification process when an outbreak occurs.
 Equipment is spread throughout the facility and may be processed or stored in numerous locations, making it difficult to track the equipment for documentation purposes.

Adapted from The Joint Commission.²

Hamm et al. Reducing the risk of transmitting infection by transvaginal ultrasound examination. *Am J Obstet Gynecol* 2020.

facilities—rates that have steadily increased since 2009. Several contributory factors were identified, including lack of appreciation for the risk of transmission, lack of staff knowledge or training, and lack of priority given by leadership to infection prevention (Box). In a 2018 survey of 358 US healthcare facilities, 20% reported incorrect processing of ultrasound probes, with most of those cases in obstetrics and gynecology and emergency departments.³

Are sterile probe covers sufficient to prevent transmission of infection?

Some commercially produced ultrasound probe covers have unacceptably high rates of leakage (8% to 81%)^{20,21} and cannot realistically be considered to offer an effective barrier against transmission of infection, particularly viruses. There are at least 2 commercial probe covers that are considered to be “viral barriers” and have demonstrated a low failure rate for leakage of particles as small as 20 nm in laboratory tests.

Alternatives to commercial probe covers include condoms and surgical gloves. These items are manufactured to meet strict Acceptable Quality Limit (AQL) standards that reflect the percentage of items in a sample that are free from holes. The AQL is 0.25% for condoms, 1.5% for surgical gloves, and 2.5% for standard examination gloves.^{1,22,23} Leakage rates of 1% to 2% have been reported when condoms are used to cover endovaginal ultrasound probes.^{20,24} In a study of 500 commercially produced probe covers of 10 brands, leakage rates ranged from 0% to 5%.²⁵

Recommendations: Disposable, single-use “viral barrier” probe covers or condoms are recommended for transvaginal ultrasound examination because of their relatively low failure rate. Standard examination gloves and commercially produced probe covers without proven viral

protection are not recommended. However, all probe covers have a finite failure rate and cannot be relied on as the sole method of infection prevention. Thus, the ultrasound probe itself must be cleaned and disinfected after each transvaginal ultrasound examination.

Is it sufficient to wipe down the probe and spray it with a disinfectant after each use?

Wiping the probe to remove ultrasound coupling gel and other visible debris is a component of cleaning, as defined in Table 1. Additional cleaning may include rinsing the probe under running water and then drying thoroughly with a soft, clean towel. Disinfectant sprays such as phenolics or

TABLE 1
Levels of cleaning, disinfection, and sterilization

Level	Definition
Cleaning	Removal of visible material from objects and surfaces, usually accomplished manually or mechanically using water with detergents or enzymatic products
Low-level disinfection	Destruction of most bacteria, some viruses, and some fungi, but does not necessarily inactivate <i>Mycobacterium tuberculosis</i>
Mid-level disinfection	Inactivation of <i>M. tuberculosis</i> , bacteria, most viruses, most fungi, and some bacterial spores
High-level disinfection	Destruction and/or removal of all microorganisms except bacterial spores
Sterilization	Destruction and/or elimination of all forms of microbial life

Adapted from Rutala et al.²⁶

Hamm et al. Reducing the risk of transmitting infection by transvaginal ultrasound examination. *Am J Obstet Gynecol* 2020.

TABLE 2
Recommended disinfection levels for medical devices

Type of device	Definition	Examples used in obstetrics and gynecology	Minimum processing recommended after each use
Critical	Objects that enter sterile tissues or the vascular system, resulting in a high risk for infection if contaminated with any microorganism	Surgical instruments, laparoscopes, implants, intraoperative ultrasound probes	Cleaning, then sterilization
Semicritical	Items that contact mucus membranes or nonintact skin	Vaginal ultrasound probes, abdominal ultrasound probes used on nonintact skin, vaginal speculums, diaphragm fitting rings	Cleaning, then high-level disinfection
Noncritical	Items that only contact intact skin	Abdominal ultrasound probes, stethoscopes	Cleaning as needed, then low-level disinfection

Adapted from Rutala et al.²⁶

Hamm et al. Reducing the risk of transmitting infection by transvaginal ultrasound examination. *Am J Obstet Gynecol* 2020.

quaternary ammonium are low-level disinfectants.²⁶ These are generally acceptable for transabdominal ultrasound probes and other external-use devices, which are classified as “noncritical” devices, as shown in Table 2. However, low-level disinfection is insufficient for transvaginal ultrasound probes and other internal-use devices because methicillin-resistant *Staphylococcus* and other pathogenic organisms can survive such treatment.^{13,14,27} Transvaginal ultrasound probes are considered “semi-critical” devices and are recommended to undergo high-level disinfection after each use.²⁶

Recommendation: After each use, transvaginal ultrasound probes should be cleaned and then treated with high-level disinfection. Wipe down and spray are low-level disinfection procedures and are not sufficient for transvaginal probes.

Are glutaraldehyde and orthophthalaldehyde adequate for high-level disinfection of vaginal ultrasound probes?

Table 3 shows the agents considered to be high-level disinfectants by the US Food and Drug Administration (FDA). Over 2 dozen products have aldehydes as an active ingredient. Unfortunately, in bench-testing, these aldehydes have virtually no viricidal activity against HPV.^{16,17,28} Phenol and isopropyl alcohol also fail to neutralize HPV.¹⁶ In US women aged 18 to 59 years, the prevalence of genital HPV is 49.9%, and the prevalence of high-risk genital HPV is 20.4%.^{18,19} Thus, if the probe cover fails during a transvaginal ultrasound examination, there is a high probability that the probe will become contaminated with HPV. If the disinfectant solution is ineffective against HPV, there is a high probability that the virus will persist on the probe until the next examination. Thus, a transvaginal ultrasound examination may potentially expose an individual to high-risk HPV, the virus that causes most cases of cervical cancer.^{18,19}

The high-level disinfection agents that have been shown to be effective against HPV in bench-testing include hydrogen peroxide (sonicated system),¹⁷ hypochlorite,^{16,28} and peracetic acid.¹⁶ We are not aware of bench-testing to evaluate whether soaking systems based on hydrogen peroxide, with or without peracetic acid, have anti-HPV viricidal activity.

Ultraviolet C (UVC) radiation is also effective against HPV.²⁸ A UVC chamber device is approved by the FDA for low-level disinfection.^{29,30} Bench-testing of the device showed reasonable efficacy against pathogenic vaginal bacteria but persistent contamination by other bacteria.³¹

TABLE 3
High-level disinfectants and chemical sterilants cleared by the US Food and Drug Administration

Active ingredients	Number of products listed	Efficacy against HPV
Glutaraldehyde	16	No
Glutaraldehyde + isopropanol	3	No
Glutaraldehyde + phenol or phenate	1	No
Hydrogen peroxide	3	Yes
Hydrogen peroxide + peracetic acid	2	?
Hypochlorite + hypochlorous acid	2	Yes
Orthophthalaldehyde	6	No
Peracetic acid	3	Yes

HPV, human papillomavirus.

Adapted from US Food and Drug Association.²⁹ The source lists 36 specific products, with varying concentrations of ingredients and specific time and temperature requirements for demonstrated efficacy.

Hamm et al. Reducing the risk of transmitting infection by transvaginal ultrasound examination. *Am J Obstet Gynecol* 2020.

Thus, the UVC chamber alone is not considered sufficient for the disinfection of vaginal ultrasound probes.

Recommendations: After each use, the transvaginal ultrasound probe should undergo high-level disinfection using 1 of the agents with proven efficacy against HPV, such as hydrogen peroxide (sonicated system), hypochlorite, or peracetic acid. Each probe manufacturer has a list of agents compatible with its probes. Disinfection procedures should follow the time and temperature recommendations published by the chemical manufacturer. Because glutaraldehyde, orthophthalaldehyde, phenols, and isopropyl alcohol are ineffective against HPV, if these agents are used for high-level disinfection, the probe should also be treated using a UVC chamber disinfection device with proven efficacy against HPV.

Are precautions needed to prevent transmission of infection via ultrasound coupling gel?

Multiple-use gel containers can potentially harbor pathogenic organisms, especially if they are kept in a warming device.^{14,31}

Recommendations: If multiple-use containers are used for transabdominal ultrasound examination, the containers should be discarded when empty and should not be refilled. The use of gel from single-use, sterile, disposable gel packets for both inside and outside the probe cover is recommended to minimize the risk of contamination during transvaginal ultrasound examinations.¹ However, there have been outbreaks of bacterial infection caused by contaminated packets of gel labeled as “sterile” by the manufacturer.^{32–35}

Summary and conclusions

The following steps are recommended to minimize the risk of transmitting vaginal infections via transvaginal ultrasound examination:

1. During each examination, the probe should be covered with a sterile, disposable, single-use cover with a documented low rate of leakage—either a “viral barrier” probe cover or a condom.
2. The ultrasound coupling gel should be from sterile, disposable, single-use packets.
3. After each examination, the probe should be cleaned to remove any gel or visible debris.
4. After cleaning, the probe should undergo high-level disinfection using a method with proven efficacy against bacteria and viruses, including HPV, such as hydrogen peroxide (sonicated system), hypochlorite, or peracetic acid. Glutaraldehyde, orthophthalaldehyde, phenol, and isopropyl alcohol are not effective against HPV.

Each step alone is insufficient to prevent transmission of infection. The safest approach is to apply all these steps together.³⁶

In light of the reasons for noncompliance with infection prevention standards noted by The Joint Commission and summarized in the [Box](#), an effective program for infection prevention requires a commitment on the part of the leadership of each practice or facility, designation of key personnel who will be responsible for infection prevention, development of written procedures, and training and monitoring of the ultrasonography staff.

Additional information about cleaning and disinfection of ultrasound equipment and other medical devices can be found in excellent overviews by the American Institute of Ultrasound in Medicine¹ and the Centers for Disease Control and Prevention,²⁶ respectively. The latter document also addresses environmental safety, handling, and disposal of chemical disinfectants—topics that are beyond the scope of this review. ■

REFERENCES

1. American Institute of Ultrasound in Medicine. Guidelines for cleaning and preparing external- and internal-use ultrasound transducers between patients, safe handling, and use of ultrasound coupling gel. AIUM official statement. 2018. Available at: https://www.aium.org/accrreditation/Guidelines_Cleaning_Preparing.pdf. Accessed August 5, 2020.
2. The Joint Commission. Improperly sterilized or HLD equipment—a growing problem. Quick Safety. 2017. Available from: https://www.jointcommission.org/assets/1/23/qs_33a_2017.pdf. Accessed August 5, 2020.
3. Carrico RM, Furmanek S, English C. Ultrasound probe use and reprocessing: results from a national survey among U.S. infection preventionists. *Am J Infect Control* 2018;46:913–20.
4. Scibetta EW, Han CS. Ultrasound in early pregnancy: viability, unknown locations, and ectopic pregnancies. *Obstet Gynecol Clin North Am* 2019;46:783–95.
5. Mei JY, Afshar Y, Platt LD. First-trimester ultrasound. *Obstet Gynecol Clin North Am* 2019;46:829–52.
6. Friedman AM, Srinivas SK, Parry S, Elovitz MA, Wang E, Schwartz N. Can transabdominal ultrasound be used as a screening test for short cervical length? *Am J Obstet Gynecol* 2013;208:190.e1–7.
7. Pandipati S, Combs CA, Fishman A, Lee SY, Mallory K, Ivanovich F. Prospective evaluation of a protocol for using transabdominal ultrasound to screen for short cervix. *Am J Obstet Gynecol* 2015;213:99.e1–13.
8. Werner EF, Hamel MS, Orzechowski K, Berghella V, Thung SF. Cost-effectiveness of transvaginal ultrasound cervical length screening in singletons without a prior preterm birth: an update. *Am J Obstet Gynecol* 2015;213:554.e1–6.
9. Campbell S. Universal cervical-length screening and vaginal progesterone prevents early preterm births, reduces neonatal morbidity and is cost saving: doing nothing is no longer an option. *Ultrasound Obstet Gynecol* 2011;38:1–9.
10. Combs CA. Vaginal progesterone for asymptomatic cervical shortening and the case for universal screening of cervical length. *Am J Obstet Gynecol* 2012;206:101–3.
11. Khalifeh A, Berghella V. Universal cervical length screening in singleton gestations without a previous preterm birth: ten reasons why it should be implemented. *Am J Obstet Gynecol* 2016;214:603.e1–5.
12. Society for Maternal-Fetal Medicine Publications Committee, with assistance of Vincenzo Berghella. Progesterone and preterm birth prevention: translating clinical trials data into clinical practice. *Am J Obstet Gynecol* 2012;206:376–86.
13. Casalegno JS, Le Bail Carval K, Eibach D, et al. High risk HPV contamination of endocavity vaginal ultrasound probes: an underestimated route of nosocomial infection? *PLoS One* 2012;7:e48137.

14. Westerway SC, Basseal JM, Brockway A, Hyett JA, Carter DA. Potential infection control risks associated with ultrasound equipment—a bacterial perspective. *Ultrasound Med Biol* 2017;43:421–6.
15. M'zali F, Bounizra C, Leroy S, Mekki Y, Quentin-Noury C, Kann M. Persistence of microbial contamination on transvaginal ultrasound probes despite low-level disinfection procedure. *PLoS One* 2014;9:e93368.
16. Meyers J, Ryndock E, Conway MJ, Meyers C, Robison R. Susceptibility of high-risk human papillomavirus type 16 to clinical disinfectants. *J Antimicrob Chemother* 2014;69:1546–50.
17. Ryndock E, Robison R, Meyers C. Susceptibility of HPV16 and 18 to high level disinfectants indicated for semi-critical ultrasound probes. *J Med Virol* 2016;88:1076–80.
18. Bosch FX, Manos MM, Muñoz N, et al. Prevalence of human papillomavirus in cervical cancer: a worldwide perspective. International biological study on cervical cancer (IBSCC) study group. *J Natl Cancer Inst* 1995;87:796–802.
19. Muñoz N, Bosch FX, De Sanjosé S, Shah KV. The role of HPV in the etiology of cervical cancer. *Mutat Res* 1994;305:293–301.
20. Rooks VJ, Yancey MK, Elg SA, Brueske L. Comparison of probe sheaths for endovaginal sonography. *Obstet Gynecol* 1996;87:27–9.
21. Hignett M, Claman P. High rates of perforation are found in endovaginal ultrasound probe covers before and after oocyte retrieval for in vitro fertilization-embryo transfer. *J Assist Reprod Genet* 1995;12:606–9.
22. Male latex condom: specification, prequalification and guidelines for procurement, 2010. World Health Organization. 2010. Available at: https://www.who.int/reproductivehealth/publications/family_planning/9789241599900/en/. Accessed August 5, 2020.
23. Patient examination gloves and surgeon's gloves; sample plans and test method for leakage defects; adulteration. Fed Regist 2011. To be codified at 21 CFR §800.20. Available at: <https://www.govinfo.gov/app/details/CFR-2011-title21-vol8/CFR-2011-title21-vol8-sec800-20>. Accessed August 5, 2020.
24. Milki AA, Fisch JD. Vaginal ultrasound probe cover leakage: implications for patient care. *Fertil Steril* 1998;69:409–11.
25. Basseal JM, Westerway SC, Hyett JA. Analysis of the integrity of ultrasound probe covers used for transvaginal examinations. *Infect Dis Health* 2020;25:77–81.
26. Rutala WA, Weber DJ. Healthcare Infection Control Practices Advisory Committee. Guideline for disinfection and sterilization in healthcare facilities, 2008. 2008. Available at: <https://www.cdc.gov/infectioncontrol/pdf/guidelines/disinfection-guidelines-H.pdf>. Accessed August 5, 2020.
27. Ngu A, McNally G, Patel D, Gorgis V, Leroy S, Burdach J. Reducing transmission risk through high-level disinfection of transvaginal ultrasound transducer handles. *Infect Control Hosp Epidemiol* 2015;36:581–4.
28. Meyers C, Milici J, Robison R. UVC radiation as an effective disinfectant method to inactivate human papillomaviruses. *PLoS One* 2017;12:e0187377.
29. US Food and Drug Administration. FDA-cleared sterilants and high level disinfectants with general claims for processing reusable medical and dental devices. 2019. Available at: <https://www.fda.gov/medical-devices/reprocessing-reusable-medical-devices-information-manufacturers/fda-cleared-sterilants-and-high-level-disinfectants-general-claims-processing-reusable-medical-and>. Accessed August 5, 2020.
30. Ultraviolet (UV) radiation chamber disinfection device. Fed Regist 2019. To be codified at 21 CFR §880.6600. Available at: <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfcfr/CFRSearch.cfm?CFRPart=880&showFR=1&subpartNode=21:8.0.1.1.27.6>. Accessed August 5, 2020.
31. Schmitz J, Kossow A, de Murcia KO, et al. Disinfection of transvaginal ultrasound probes by ultraviolet C—a clinical evaluation of automated and manual reprocessing methods. *Ultraschall Med* 2019 [Epub ahead of print].
32. Ricci S, Pinette MG, Wax JR, Craig W, Forrest L, Dragoni C. The effect of temperature on bacterial growth in the presence of nonsterile ultrasound coupling gel. *Am J Obstet Gynecol* 2020;222:188.
33. Solaimalai D, Ragupathi NKD, Ranjini K, et al. Ultrasound gel as a source of hospital outbreaks: Indian experience and literature review. *Indian J Med Microbiol* 2019;37:263–7.
34. Provenzano DA, Liebert MA, Steen B, Lovetro D, Somers DL. Investigation of current infection-control practices for ultrasound coupling gel: a survey, microbiological analysis, and examination of practice patterns. *Reg Anesth Pain Med* 2013;38:415–24.
35. Shaban RZ, Maloney S, Gerrard J, et al. Outbreak of health care-associated Burkholderia cenocepacia bacteremia and infection attributed to contaminated sterile gel used for central line insertion under ultrasound guidance and other procedures. *Am J Infect Control* 2017;45:954–8.
36. Combs CA, Fishman A. A proposal to reduce the risk of transmission of human papilloma virus via transvaginal ultrasound. *Am J Obstet Gynecol* 2016;215:63–7.

From the Patient Safety and Quality Improvement Committee, Society for Maternal-Fetal Medicine, Washington, DC.

The authors report no conflict of interest.

All authors and Committee members have filed a conflict of interest disclosure delineating personal, professional, and/or business interests that might be perceived as a real or potential conflict of interest in relation to this publication. Any conflicts have been resolved through a process approved by the Executive Board. The Society for Maternal-Fetal Medicine (SMFM) has neither solicited nor accepted any commercial involvement in the development of the content of this publication.

This document has undergone an internal peer review through a multilevel committee process within SMFM. This review involved critique and feedback from the SMFM Publications and Document Review Committees and final approval by the SMFM Executive Committee. SMFM accepts sole responsibility for the content of the document. SMFM publications do not undergo editorial and peer review by the American Journal of Obstetrics and Gynecology. The SMFM Patient Safety and Quality Committee reviews publications every 36 to 48 months and issues updates as needed. Further details regarding SMFM publications can be found at www.smfm.org/publications.

SMFM has adopted the use of the word “woman” (and the pronouns “she” and “her”) to apply to individuals who are assigned female sex at birth, including individuals who identify as men as well as nonbinary individuals who identify as both genders or neither gender. As gender-neutral language continues to evolve in the scientific and medical communities, SMFM will reassess this usage and make appropriate adjustments as necessary.

All questions or comments regarding the document should be referred to the SMFM Patient Safety and Quality Improvement Committee at smfm@smfm.org.